

349KLX

FAQs

Does the 349KLX have a unit lock out feature?

Locking the Weight Measuring Unit

The weight measuring unit can be locked to only measure in KG or LB. Follow the procedures below to lock the weight measuring unit.

KG Lock Instructions

1. Press the **ON** button to turn the scale on. Wait for "0.0" to appear on the display.
2. Press and hold the **lb./kg** button so that KG icon is displayed and continue holding the **lb./kg** button for approximately 10 seconds.
3. When the KG icon flashes, release the **lb./kg** button. The KG weight measuring unit is now locked.

LB Lock Instructions

Note: To lock in LB weight mode, you must follow the instructions above to lock in KG weight mode first.

1. Once the scale is locked in KG mode, press and hold the **lb./kg** button for approximately 10 seconds until the LB icon flashes.
2. When the LB icon flashes, release the **lb./kg** button. The LB weight measuring unit is now locked.

Unlocking the Weight Measuring Unit

Note: If your scale is locked in KG mode, you must first follow the instructions above to lock into LB mode, and then begin the disabling instructions below.

1. Press the **ON** button to turn the scale on. Wait for "0.0" to appear on the display.
2. Press and hold the **lb./kg** button until LB and KG icons both flash. When LB and KG are flashing, release the **lb./kg** button. The weight measuring unit is now unlocked; pressing the **lb./kg** button will allow user to toggle between LB and KG.

What type of batteries does the 349KLX use?

The current version of the 349KLX uses six AA batteries. Models with a date code of DC5016 and earlier use one 9V battery.

Can the display head on a 349KLX be replaced?

The display head on models with a date code with the last two digits higher than 16 can be replaced. To purchase a replacement display head, contact your medical supply distributor and order item # DISPLAY349. The display head on models with a date code lower than DC5016 cannot be replaced.

Is there a replacement power adapter for the 349KLX?

The power adapter on models with a date code with the last two digits higher than 07 can be replaced. To purchase a replacement power adapter, contact your medical supply distributor and order item #ADPT40.

How much weight is needed to calibrate the 349KLX?

400 pounds of certified weights are required and employed in 50 / 100 / 200 / 400 lbs increments.

What are the platform dimensions of the 349KLX?

12 1/2" (w) x 12" (d) x 1 7/8" (h)/318mm (w) x 305mm (d) x 48mm (h)

Does the 349KLX have an auto off feature?

Auto-Off Function

As an energy saving feature, the scale automatically powers off after 2 minutes of inactivity. This function is enabled by default.

Disable Auto-Off

1. While the scale is powered off, press and hold the **lb./kg** button.
2. While holding the **lb./kg** button, press the **ON** button to power the scale on.
3. "349" will briefly appear on the display followed by "A-OF" and then "0.0" lb appears.
4. Release the **lb./kg** and **ON** button.
5. Auto-off is now disabled, and the scale will remain on until the scale is manually powered off

Enable Auto-Off

1. While the scale is powered off, press and hold the **lb./kg** button.
2. While holding the **lb./kg** button, press the **ON** button to power the scale on.
3. "349" will briefly appear on the display, followed by "0.0" lb.
4. Release the **lb./kg** and **ON** button.
5. Auto-off is now enabled.

Does the 349KLX have a low battery indicator?

The 349KLX will display the word "LO" when it is time to replace the batteries.

What is the warranty on the 349KLX?

The warranty on the 349KLX is a 2 year manufacturer limited warranty.