


# Certificate

**Polymer Technology Systems Inc. (PTS Diagnostics)**

using

**A1CNow+, A1CNow SELF CHECK**

participated in the IFCC HbA1c Certification Programme to demonstrate traceability to the IFCC Reference Measurement Procedure and performed as shown below.




Total Error =	4.0	mmol/mol
Bias =	-1.1	mmol/mol
Imprecision =	2.9	%
Grade =	Standard	

Criteria derived from the IFCC model for Quality Targets HbA1c (Clin Chem 2015;61 : 752-59)

Date of Certification : 01 January 2024

Date of Expiry : 01 January 2025



IFCC Network Coordinator  
C. Siebelder